

Steris Steriliser

Activity		Type Of Service				
		A	B	C	D	E
Last Issue Date: 08/09/2014						
1	Discuss equipment performance with operators and inspect printouts for any history of failures.			Y		
2	Record previous diagnostic cycle variables if available.			Y		
3	Test printer.			Y		
4	Test inflatable seal and air system for leakage.			Y		
5	Check and clean chamber and lid seal.			Y		
6	Ensure lid latch bolt is secure.			Y		
7	Inspect power cord for any damage or fraying.			Y		
8	Clean air vents in electronics enclosure.			Y		
9	Lubricate lid latches & check gas spring cylinder.			Y		
10	If applicable check resistance of switches on switch panel and replace if necessary. (for Model 90A1 replace if resistance above 80 ohms).			Y		
11	Calibrate conductivity monitor.			Y		
12	Calibrate pressure transducer.			Y		
13	Check operation of pump including its pressure switch.			Y		
14	Check filters.			Y		
15	Run diagnostic cycle.			Y		
16	Run sterile cycle.			Y		
17	With site approval undertake any repairs required.			Y		
18	Advice the site in writing if the unit can not achieve 100% efficiency.			Y		
19	Undertake Calibration.			Y		
20	Undertake Validation.				Y	
21	Undertake independent inspection as required for the Registration.					Y
22	Register unit.					Y
23	Record results in log book.			Y		

**Special Comments and Technical Data**

C SERVICE 6 MONTHLY  
 D SERVICE ANNUAL  
 E INDEPENDENT INSPECTION FREQUENCY - EVERY 2 YEARS  
 E REGISTRATION FREQUENCY - EVERY 5 YEARS

**Disclaimer** -This TDS is intended to provide guidance only to support Preventative Maintenance servicing activities. All TDS users are encouraged to read and understand the full [Conditions of Use](#) provided on the final page of this document and on the DPTI website.

For more information please contact the applicable Facilities Management Service Provider.



Government of South Australia  
 Department of Planning,  
 Transport and Infrastructure

Download Date: 22/12/2020

# Conditions of Use

A TDS should be used as a generic guide for Facilities Management Service Providers (FMSP) to deliver Facilities Maintenance Services, specifically Preventative Maintenance in accordance with the DPTI [Agency Work Procedure Manual](#).

Users of a TDS should familiarise themselves with the following **Conditions of Use**:

1. You use the information and data contained in any TDS published by DPTI at your sole risk. DPTI-AGFMA does not have the expertise to provide complete or accurate technical data schedules and provides these technical data sheets merely as a starting point and/or guide.
2. If you rely on the information in a TDS you are responsible for ensuring, by independent verification, its accuracy, currency and completeness. DPTI cannot guarantee that the information contained in a TDS meets the standards or requirements of legislative requirements.
3. A TDS is not an exhaustive list of tasks or obligations that may be required and is generally generic in nature. DPTI does not oblige the user of a TDS to rely on it to the exclusion of other sources of information. For example, manufacturer's requirements may need to be followed for additional and / or alternative tasks and for additional and / or alternative servicing frequencies. You should always check the primary source of information such as the Australian Standards, manufacturer's specifications, legislation and other standards before undertaking any work to which a schedule may apply.
4. You must apply appropriate risk management principles and carry out all tasks in accordance with obligations under the [Work Health and Safety Act 2012](#).
5. You must not use a TDS unless you meet the specific competency requirements for preventative maintenance servicing works and are capable of applying the TDS to your trade or area of expertise. A TDS is only to be used by people who are specifically instructed by DPTI to use them and only for work undertaken in South Australia.
6. DPTI, its agents, instrumentalities, officers and employees make no representations, express or implied, as to the accuracy of the information, the data contained in a TDS or the suitability of a TDS for a particular purpose.
7. DPTI does not provide legal advice. DPTI accepts no liability, howsoever arising, for any loss resulting from the use of a TDS and any information or data or reliance placed on them.
8. DPTI may change information and data in a TDS without notice.
9. DPTI may revise this disclaimer at any time by updating these Conditions of Use.