

Malmet Bedpan Washer Disinfector (B&C services)

Last Issue Date: 07/05/2018		Type Of Service				
		A	B	C	D	E
Activity						
1	Check for steam or water leaks. Tighten union, hose clamps and glands where necessary.		Y	Y		
2	Remove and clean out sprays and bottle hooks (unscrew from inside chamber). Clean in ultrasonic cleaner if available.		Y	Y		
3	Remove level probe in steam generator tank and clean off any build-up of residue.		Y	Y		
4	Remove temperature probes in steam tank and chamber and clean off any build-up of residue.		Y	Y		
5	Visually inspect build-up of residue in steam tank, especially in areas of poor water quality.		Y	Y		
6	Check filter in the water inlet solenoid valve and clean as necessary.		Y	Y		
7	Check and tighten, if necessary, all electrical connections.		Y	Y		
8	Check the alignment of the spray nozzles for the urine bottles.		Y	Y		
9	Note: Unit is to be isolated from the electric supply prior to the following servicing:			Y		
10	For access remove: a) Top panel - remove screws; b) Side panel - push back and then lift up to remove it; and c) Front panel - remove screws from the bottom.			Y		
11	Remove three sprays, two bottle hooks and rotary nozzle from inside the chamber and: a) Check that the holes are clear and clean as necessary (hold under tap and pressurise or clean in ultrasonic cleaner). In areas with hard water or high minerals in the water supply, chemical de-scaler may be the quickest and easiest means to remove built up deposits; and b) Replace the sprays and bottle hooks.			Y		
12	Start a cycle and check that the cold water pump is working, 17 seconds of the cycle.			Y		
13	After ascertaining that the pump is working properly, check: a) The solenoid operation in the cold water tank; and b) Check the steam sensor (steam generator maintenance).			Y		
14	Make sure the solenoids are completely shutting off and levels are not creeping up, if so, clean and/or replace the solenoid.			Y		
15	Tighten the screws retaining the pan rack on the door.			Y		
16	Undertake a visual inspection and if necessary, clean dust and grit off the components. Refix panels.			Y		
17	Record the results in the logbook.		Y	Y		

**Disclaimer** - This TDS is intended to provide guidance only to support Preventative Maintenance servicing activities. All TDS users are encouraged to read and understand the full [Conditions of Use](#) provided on the final page of this document and on the DPTI website.

For more information please contact the applicable Facilities Management Service Provider.



Government of South Australia  
Department of Planning,  
Transport and Infrastructure

Download Date: 22/12/2020

## Special Comments and Technical Data

---

B SERVICE 2 MONTHLY

C SERVICE 4 MONTHLY

THE MALMET MACHINE DETERGENT IS HAZARDOUS WITH A CLASS RATING OF 8, TO BE REPLENISHED BY THE FS TRADE/TECHNICIAN. FOR DETAILED SERVICING INFORMATION REFER TO THE MANUAL.

THIS TDS IS A SPLIT VERSION, FOR THE OTHER 2 SPLITS REFER:

- TDS EQ303E (BY SITE STAFF).

- TDS EQ303A (BY FS TRADE).

FOR THE PRIMARTY TDS REFER TO EQ303.

**Disclaimer** -This TDS is intended to provide guidance only to support Preventative Maintenance servicing activities. All TDS users are encouraged to read and understand the full [Conditions of Use](#) provided on the final page of this document and on the DPTI website.

*For more information please contact the applicable Facilities Management Service Provider.*



**Government of South Australia**  
Department of Planning,  
Transport and Infrastructure

**Download Date:** 22/12/2020

2 of 3

# Conditions of Use

A TDS should be used as a generic guide for Facilities Management Service Providers (FMSP) to deliver Facilities Maintenance Services, specifically Preventative Maintenance in accordance with the DPTI [Agency Work Procedure Manual](#).

Users of a TDS should familiarise themselves with the following **Conditions of Use**:

1. You use the information and data contained in any TDS published by DPTI at your sole risk. DPTI-AGFMA does not have the expertise to provide complete or accurate technical data schedules and provides these technical data sheets merely as a starting point and/or guide.
2. If you rely on the information in a TDS you are responsible for ensuring, by independent verification, its accuracy, currency and completeness. DPTI cannot guarantee that the information contained in a TDS meets the standards or requirements of legislative requirements.
3. A TDS is not an exhaustive list of tasks or obligations that may be required and is generally generic in nature. DPTI does not oblige the user of a TDS to rely on it to the exclusion of other sources of information. For example, manufacturer's requirements may need to be followed for additional and / or alternative tasks and for additional and / or alternative servicing frequencies. You should always check the primary source of information such as the Australian Standards, manufacturer's specifications, legislation and other standards before undertaking any work to which a schedule may apply.
4. You must apply appropriate risk management principles and carry out all tasks in accordance with obligations under the [Work Health and Safety Act 2012](#).
5. You must not use a TDS unless you meet the specific competency requirements for preventative maintenance servicing works and are capable of applying the TDS to your trade or area of expertise. A TDS is only to be used by people who are specifically instructed by DPTI to use them and only for work undertaken in South Australia.
6. DPTI, its agents, instrumentalities, officers and employees make no representations, express or implied, as to the accuracy of the information, the data contained in a TDS or the suitability of a TDS for a particular purpose.
7. DPTI does not provide legal advice. DPTI accepts no liability, howsoever arising, for any loss resulting from the use of a TDS and any information or data or reliance placed on them.
8. DPTI may change information and data in a TDS without notice.
9. DPTI may revise this disclaimer at any time by updating these Conditions of Use.